

## Editorial

# Coagulation Disorders Beyond Technicalities

Munawar Hussain\*, Hans P Wendel and Frank K Gehring

Department of Cardiac and Vascular Surgery, Tuebingen University, Germany

\*Corresponding author: Munawar Hussain, Biosensor Research Group, Thoracic, Cardiac and Vascular Surgery Department, Tuebingen University, Germany

Received: July 25, 2017; Accepted: July 27, 2017;

Published: August 10, 2017

## Editorial

The coagulation disorders have caught multidisciplinary attraction by the scientific community in the recent decades. There are several technicalities that hinder the pace of research. The research area demands highly specialized personals, proper technologies, techniques and methodologies. The conventional technologies and methods produce an extremely fewer pieces of information of issues and the research area suffers an infancy stage. Rationally speaking, the domain requires highly information oriented techniques to find the solutions to the clinical problems.

Majority of the anticoagulants in today's clinical laboratory inhibit the action of Factor Xa to certain extent. The determination of Factor Xa is crucial in routine clinical settings for practice and laboratories. The Xa monitoring keeps the fourth position in the standard assays, while the first three places are positioned by thromboplastin time (PT) [1-3] activated Partial Thromboplastin Time (aPTT) [4-6] and Thrombine Time (TT) [7] tests respectively. In reality, the test is not available in the hospitals and labourites everywhere even in developed countries. This is due to its tiresome nature, costly set ups and need of qualified or specialized staff.

PT is the first priority in clinical settings of innate coagulation disorders, dialysis, extensive surgery, or drugs having an inflammatory potency. For achieving an overview of the haemostatic system of patients, PT tests are combined with the tests for aPTT. PT assays give the information of the extrinsic part of the plasmatic coagulation system for instance coagulation Factor VII, and Factor X, V, II, while aPTT tests predict an isolated or a combined lack of the coagulation Factors of the intrinsic haemostatic cascade. The clinics suffer the deficiency of precise and accurate tests for monitoring of anticoagulants such as 'heparin' [8] or 'argatroban' [9] in human blood or plasma in surgical events. After one century since the first haemostatic test, clinics still require qualified personals for the PT or aPTT tests. "Prothrombinase induced Clotting Time" (PiCT) could be an alternative of aPTT, but it has not been recognized worldwide yet [10,11].

Research on thrombocytopenia is a key domain for the patients under the practice of anticoagulant therapy, but the area suffers challenges. For instance, the time-consuming and laborious step is the platelet isolation from the healthy donor's Platelet Rich Plasma (PRP) in Heparin Induced Thrombocytopenia (HIT) diagnosis



Figure 1: Exemplary acoustic (QCM-D) instrument (www.3t-analytik.de).

tests. The laboratories for the HIT tests require highly experienced personals [12,13]. Every clinical case of HIT put a average of 9000 Euro of extra costs because of extended hospital stay by a patient and the requirement of costly alternative anticoagulants (Figure 1) [14].

Modern technologies and equipment for the haemostasis studies is a sign of hope for coping the technicalities and barriers. The equipments for supporting the Point Of Care (POC) settings could reduce the technical flaws of the routine haemostatic tests. For instance, acoustic technology has potential [15] due to its cost-effectiveness and uniqueness [16-18]. The technique keeps preminent potential by applying the molecular imprinting, [19] biomedical engineering, miniaturization and sensor arrays [20].

## References

- Hussain M, Sinn S, Zeilinger M, Northoff H, Lieberzeit PA, Gehring FK. Blood coagulation thromboplastin time measurements on a nanoparticle coated quartz crystal microbalance biosensor in excellent agreement with standard clinical methods. *J Biosens. Bioelectron.* 2013; 4: 139.
- Hussain M, Zeilinger M, Northoff H, Lieberzeit PA, Gehring FK. Affinity based nanoparticles for quartz crystal microbalances sensors for thromboplastin time of human whole blood. *J Biosens Bioelectron.* 2013; 4.
- Hussain M. Prothrombin Time (PT) for human plasma on QCM-D technique: A better alternative to 'gold standard'. *UK J Pharm Biosci.* 2015; 3: 1-8.
- Hussain M. aPTT: 1<sup>st</sup> recognition for human whole blood on QCM-D technique. *UK J. Pharm. Biosci.* 2015; 3: 49-55.
- Hussain M, Northoff H, Gehring FK. QCM-D providing new horizon in the domain of sensitivity range and information for haemostasis of human plasma. *Biosens Bioelectron.* 2015; 66: 579-584.
- Hussain M, Northoff H, Gehring, FK. QCM-D beating the standard coagulometer in the domain of sensitivity range and information for hemostasis of human plasma. *J Biotechnol Biomater.* 2014; 3.
- Hussain M. Shortened 'thrombin time' monitoring on QCM-D: A better substitute of 'gold standard'. *UK J Pharm Biosci.* 2016; 4: 20-26.
- Hussain M. Ultra-sensitive detection of heparin via aPTT using plastic antibodies on QCM-D technique. *RSC Adv.* 2015; 5: 54963-54970.
- Hussain M. 'Argatroban' monitoring in human plasma: aPTT and PiCT studies on QCM-D vs 'gold standard'. *UK J Pharm Biosci.* 2015; 3: 42-48.
- Hussain M. PiCT: 1<sup>st</sup> recognition for human whole blood on QCM-D technique. *UK J Pharm Biosci.* 2015; 3: 1-8.
- Hussain M. A simultaneous monitoring of coagulation time and fibrinogen via PiCT on QCM-D. *UK J Pharm Biosci.* 2016; 4: 27-25.
- Hussain M, Gehring, FK, Sinn S, Northoff H. A straightforward detection of HIT type II via QCM-D. *UK J Pharm Biosci.* 2015; 3: 18-29.

13. Hussain M, Northoff H, Gehring FK. Detection of HIT antibody dependent platelet aggregation using novel surface imprinting approach. *Talanta*. 2016; 147: 1-7.
14. Wilke T, Tesch S, Scholz A, Kohlmann T, Greinacher A. The costs of heparin-induced thrombocytopenia: a patient-based cost of illness analysis. *J Thromb Haemost*. 2009; 7: 766-773.
15. Hussain M, Wackerlig J, Lieberzeit PA. Biomimetic strategies for sensing biological species. *Biosensors*. 2013; 3: 89-107.
16. Braun P, Drechsel H, Sterck A, Zhang J, Prepens G, Reiner T, et al. Quarzkristall-Mikrowaagen-Technologie als neue bioanalytische Plattform. *Biospektrum*. 2016; 22: 284-286.
17. Müller L, Sinn S, Drechsel H, Ziegler C, Wendel HP, Northoff H, et al. Investigation of prothrombin time in human whole-blood samples with a quartz crystal biosensor. *Anal Chem*. 2010; 82: 658-663.
18. Sinn S, Müller L, Drechsel H, Wandel M, Northoff H, Ziemer G, et al. Platelet aggregation monitoring with a newly developed quartz crystal microbalance system as an alternative to optical platelet aggregometry. *Analyst*. 2010; 135: 2930-2938.
19. Hussain M. Molecular imprinting' as multidisciplinary material science: Today and tomorrow. *Int J Adv Mater Res*. 2015; 1: 132-154.
20. Hussain M. QCM-D for haemostasis: Current status and future: a review. *UK J Pharm Biosci*. 2016; 4: 121-132.